



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/506,790	09/07/2004	Alain Jordan	2004_1361A	3508

35023 7590 03/27/2007
LUCE, FORWARD, HAMILTON & SCRIPPS LLP
11988 EL CAMINO REAL, SUITE 200
SAN DIEGO, CA 92130

EXAMINER

LACYK, JOHN P

ART UNIT	PAPER NUMBER
----------	--------------

3735

SHORTENED STATUTORY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE
3 MONTHS	03/27/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

Office Action Summary

Application No.

10/506,790

Applicant(s)

JORDAN ET AL.

Examiner

John P. Lacyk

Art Unit

3735

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 21 December 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-14 and 18-21 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-14 and 18-21 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|---|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date <u>12/21/06</u> . | 6) <input type="checkbox"/> Other: _____ |

Art Unit: 3735

1. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

2. Claims 1-14 and 18-22 are rejected under 35 U.S.C. 103(a) as being unpatentable over Dargent et al (6,547,801) in view of Mass et al (6,889,086) and Sayet et al (6,527,701).

Dargent et al discloses a gastric constriction device that have a moveable member that is moved by a motor to control the amount of constriction of the band. The device is controlled by an external controller. Dargent et al discloses the claimed device except for specifically stating the use of a stepper motor and the use of passive telemetry.

Sayet et al discloses a constricting member within the body to control the amount of constriction exerted by the constricting member. Sayet et al teaches that it is well known to use a stepper motor as an activating member to control the amount of constriction of the device. Therefore a modification of Dargent et al such the the motor used is specifically a stepper motor would have been obvious in view of Sayet et al which shows that stepper motors are specifically known in deivces implanted in the body to control movement of a constricting device. Mass et al teaches the use of passive telemetry is well known to communicate with and control implantable medical devices and discloses the advantages of the passive telemetry system over commonly known external telemetry systems (column 1, line 45- column 2, line 41). Therefore a modification of Dargent et al such that a passive telemetry system is substituted for the

Art Unit: 3735

remote telemetry system of Dargent et al would have been obvious in view of the teachings of Mass et al.

3. Applicant's arguments filed 12/21/06 have been fully considered but they are not persuasive. Applicant argues that none of the references teach a passive telemetry system that uses a stepper motor without processing by a microprocessor. Firstly the claims do not contain any limitations that the device does not contain a microprocessor. Claim 1, for one example only, includes a stepper motor, a moveable member, an oscillator where the external controller that uses passive telemetry to operate the device. As can be seen in the rejection Dargent et al discloses the device except for specifically using a stepper motor and passive telemetry. The secondary references to Mass et al and Sayet et al further teach the use of stepper motors and passive telemetry in similar devices as shown in the rejection. Therefore the rejection is still considered to be proper. Further applicant argues that neither Mass or Sayet teach actuation of a moveable member in an intracorporeal device by passive telemetry, however the test for obviousness is not whether the features of a secondary reference may be bodily incorporated into the structure of the primary reference; nor is it that the claimed invention must be expressly suggested in any one or all of the references. Rather, the test is what the combined teachings of the references would have suggested to those of ordinary skill in the art. See *In re Keller*, 642 F.2d 413, 208 USPQ 871 (CCPA 1981). Clearly, as discussed in the above rejection, the references

to Mass and Sayet teach the well known use of stepper motors and passive telemetry in devices that are implanted into the body.

4. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).


A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to John P. Lacyk whose telephone number is 571-272-4728. The examiner can normally be reached on Mon-Fri, 8:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Chuck Marmor, II can be reached on 571-272-4730. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3735

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



John P Lacyk
Primary Examiner
Art Unit 3735

J.P. Lacyk